

nd Testing of Electrical, Electronic & Precision Instruments)

agar, Co-operative Society, Besides Tadbund Hanuman Temple, Sikh Village, Secunderabad - 500 009, T.S, INDIA ☎: (040) 27840697, 48541008
 293008, 9948294008 Email: msudheer_98@yahoo.com, msudheer_98@smentp.org, SME_003@YAHOO.IN, url : www.smentp.org

CALIBRATION CERTIFICATE

CALIBRATION ON 25/05/2023	CALIBRATION DUE 24/05/2024
1. Calibration Report Ref. No. & Date :	SME/NP/SVMC/TIM/022/2023;Dt:25/05/2023
2. Requested by : (Name and address of the organization)	To, Prof & HOD (State reference Lab) Department Of Microbiology SV Medical College Tirupathi AP
3. Instrument Name:	Digital Timer
4. Instrument Make/Model:	Aceteq / DT-1045
5. Instrument ID. No:	SVMC/TIM/022
6. Range :	0 to 99 min
7. Accuracy :	+/- 1 % FSD
8. Environmental Condition	Temp/RH : 28 +/- 2 Deg C; R.H: 65 +/- 5%
9. Test parameter & condition :	Time V/S Time (Comparison Method)
10. Standard Instruments used for Calibration:	Time Interval Meter
11. Traceability Certificate Ref. No.	C-220622-5-1 ,Date: 29/06/2022 , Due Dt: 29/06/2023

TEST RESULTS

Sl No	Standard Reading Of Time Hr:min:sec:ms	Observed Reading Of Time min:sec	Error Value(+/-) Hr:min:sec:ms
1	00:01:00:41	01:00	00:00:00:41
2	00:10:00:19	10:00	00:00:00:19
3	00:30:00:47	30:00	00:00:00:47
4	01:50:00:30	90:00	00:00:00:30

PLEASE NOTE: THE INSTRUMENTS ERROR FOUND TO BE WITHIN THE SPECIFIED ACCURACY LIMIT AND THIS CERTIFICATE VALID FOR PERIOD OF ONE YEAR FROM THE DATE OF ISSUE.

CALIBRATED BY
S. M. Enterprises

ENGINEER IN CHARGE

