

**CENTRIFUGE Calibration / Verification Record Sheet**

Location: ICTC-CHC-DHONE-KURNOOL

Range : Up to 3500RPM

Equipment ID: APSACS/ICTC/CHC/DHN/CEE/01

SL.NO : 20070005

Calibration Procedure Reference: In-house verification of centrifuge using certified standard calibrated by Tachometer

Date of Calibration : 22.08.2023

Due date of Calibration:

21.08.2024

Tachometer ID : KMC/MB/HIVTL/SRL/TM/01

Calibration date : 19.04.2023

Due date of Calibration : 18.04.2024

Verification Done by: B.BHASKAR, Technical Officer, SRL, KMC,KURNOOL

S.NO	STANDARD READING(RPM)	UUC READING (RPM)	ERROR (RPM)
1.	501	500	-1.0
2.	997.5	1000	2.5
3.	1997.3	2000	2.7
4	2996.7	3000	3.3
5.	3496.1	3500	3.9

Accept/ Reject: Accepted since average difference is within +/- 100RPM

Final Interpretation: Centrifuge is in good working condition

Action Taken: - NA

Remark: - NA

*B. Bhaskar*  
13  
7.0