

Timer Calibration / Verification Record Sheet									
Location: ICTC-DHONE									
Equipment ID: APSACS/ICTC/CHC/DHN/TIMER/01									
Calibration Procedure Reference: In-house verification of timer using certified standard calibrated Timer									
Date of Calibration :		22.08.2023				Due date of Calibration:		21.08.2024	
Certified timer ID: APSACS/KMC/SRL/TIM/01					Calibration date : 05/04/2023		Due date of Calibration : 04/04/2024		
Verification Done by: B.BHASKAR, Technical Officer, SRL, KMC, KURNOOL									
Sr.No.	Timer Setting (Sec)	Certified Time (Sec)	Time Difference (Sec)	Timer Setting (Sec)	Certified Time (Sec)	Time Difference (Sec)	Timer Setting (Sec)	Certified Time (Sec)	Time Difference (Sec)
1	300	300	0	600	599	-1	900	900	0
2	300	299	-1	600	601	1	900	900	0
3	300	300	0	600	601	1	900	899	-1
4									
5									
Mean	300	299.6	-0.4	600	600.3	0.3	900	899.9	-0.1
SD	-	0.58	-	-	1.15	-	-	0.58	-
%CV	-	0.19	-	-	0.19	-	-	0.064	-
% Accuracy	-	-0.13	-	-	0.05	-	-	-0.011	-
Accept/ Reject: Accepted since average difference is within +/- 60 Second									
Final Interpretation: Timer is in good working condition									
Action Taken: - NA									
Remark: - NA									

B. Bhaskar
T/O