



**HIV Testing Laboratory**  
**Department of Microbiology**  
**Rangaraya Medical College, Kakinada**

RMC/HIVTL/Form No:02

**Proficiency Panel Result of ICTC from SRL**

Respected Sir / Madam,

Results of proficiency testing sera conducted on 9th August 2023.

Name of the ICTC & Code	Sample Code	True Status	Centre result Name of the test kit			Matching yes/no	Matching %	Remarks
			Rapid	Rapid	Rapid			
<b>Pithapuram , IPM 09/01/New Centre</b>	A	Non-Reactive	Comb AIDS Lot: 4000028036, Exp: 28.11.2024	Standard Q; Lot: C057090; Exp: 08.2024	HIV - Trispot Lot: 37230318; Lot:08.2024	Yes	100%	Nil
	B	Reactive - 1	-	-	-			
	C	Non-reactive	-	-	-			
	D	Non-reactive	-	-	-			

In-charge SRL  
*Phalguni*  
23/8/23  
State Reference Laboratory  
Dept. of Microbiology  
RMC, GGH Campus, KAKINADA.