



S.M. ENTERPRISES


27th
Years of Service

ISO 9001 : 2015
RQ91/3144

(Calibration and Testing of Electrical, Electronic & Precision Instruments)

Office/Cal. Lab : Plot No. 12, "Ganesh Niwas", Sanjeevaiah Nagar, Co-operative Society, Besides Tadburd Hanuman Temple, Sikh Village, Secunderabad - 500 009, T.S, INDIA z: (040) 27840697, 48541008
Cell: 9848020798, 9948291008, 9948292008, 9848293008, 9948294008 Email: msudheer_98@yahoo.com, msudheer_98@smentp.org. SME_003@YAHOO.IN, url : www.smentp.org

CALIBRATION CERTIFICATE

CALIBRATION ON 15/05/2023		CALIBRATION DUE 14/05/2024	
1. Calibration Report Ref. No. & Date :		SME/NP/GGH/TIM/03/2023;Dt:15/05/2023	
2. Requested by : (Name and address of the organization)		To, Prof & HOD(Department Of Microbiology), O.P NO-7,R.R Road,Govt General Hospital, Kakinada-533001	
3. Instrument Name:		Digital Timer	
4. Instrument Make/Model:		Aceteq / DT-1045	
5. Instrument ID. No:		GGH/TIM/03	
6. Range :		0 to 99 min	
7. Accuracy :		+/- 1 % FSD	
8. Environmental Condition		Temp/RH : 28 +/- 2 Deg C; R.H: 65 +/- 5%	
9. Test parameter & condition :		Time V/S Time (Comparison Method)	
10. Standard Instruments used for Calibration:		Time Interval Meter	
11. Traceability Certificate Ref. No.		C-220622-5-1 ,Date: 29/06/2022 , Due Dt: 29/06/2023	
TEST RESULTS			
SI No	Standard Reading Of Time	Observed Reading Of Time	Error Value(+/-)
	Hr:min:sec:ms	min:sec	Hr:min:sec:ms
1	00:01:00:48	01:00	00:00:00:48
2	00:10:00:38	10:00	00:00:00:38
3	00:30:00:55	30:00	00:00:00:55
4	01:50:00:39	90:00	00:00:00:39
PLEASE NOTE: THE INSTRUMENTS ERROR FOUND TO BE WITHIN THE SPECIFIED ACCURACY LIMIT ANDTHIS CERTIFICATE VALID FOR PERIOD OF ONE YEAR FROM THE DATE OF ISSUE.			
CALIBRATED BY S . M. Enterprises  ENGINEER INCHARGE			



GSTIN : 36AFVPS4924F1ZW Service Tax No. : AFVPS4924FST001 PAN No. : AFVPS4924F
NABL APPROVED LAB (CC-2471) / Approved Laboratory of ISO 9001-2015
TEST EQUIPT. / STANDARDS USED ARE TRACEABLE TO NATIONAL STANDARDS