

**CERTIFICATE OF CALIBRATION**

Customer Name	GOVERNMENT AREA HOSPITAL	Certificate No.	AHTUN030
Address	YANAMALA ST, TUNI, ANDHRA PRADESH, 533401	Date of Calibration	20-5-2023
		Valid Upto	19-05-2024
		Date of Issue	25-05-2023

DUC* Details			
DUC* Name	Centrifuge	S No.	NA
Make/Model No.	HETTICH/EBA 8	Location	Lab
ID No.	07070037	Visual Inspection	OK

Details of Reference Standards & Major Equipments Used		
Equipment Name	DIGITAL TACHO METER	THERMO HYGROMETER
Make	METRAVI	HTC-2
Model/SR No.	NCTM 1000/190518478	...
Equipment Id	CHPL/MT-KL/TAM/002	
Calibration certificate No	UCSL/07-22/184_06	UCSL/07-22/184_01
Calibration Validity	20/07/2023	20/07/2023
Calibration By	UNIQUE CALIBRATIONS	UNIQUE CALIBRATION SOLUTIONS

Environmental Condition			
Temperature	27°C	Relative Humidity	54%

Calibration Results				
SL NO	SET READING (RPM)	MEASURED READING (RPM)	TOLERANCE	STATUS
1	500	500.6	±5%	PASS
2	1000	1019	±5%	PASS
3	2500	2532	±5%	PASS

CALIBRATION STATUS	PASS	✓	FAIL	
--------------------	------	---	------	--

REMARKS:- \*DUC- DEVICE UNDER CALIBRATION

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By



Issued and Approved By  
PREJITH T B  
TECHNICALMANAGER