

## VIEW LAB MONTHLY SUMMARY

Lab Name **KOVAL DIAGNOSTIC LABORATORY**[Details About Robust Analysis](#)Lab No **5675**[Detail About Monthly Summary](#)Month **June**[Detail about SDI](#)Year **2023**Constituent Group **Chemistry II**[Click on the analyte to view Graphical Data](#)[All Analyser Result](#)[Print](#)Date of Result Entered : **13/06/2023**Date of Report Published : **04/07/2023**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	<a href="#">GLUCOSE II</a>	GOD-POD II	Any Analyser (Automation / Semi Automation )	1526	266.99	9.70	25.91	253.6 mg/dL	-0.52	1.33
2	<a href="#">UREA II</a>	Urease UV / GLDH II	Any Analyser (Automation / Semi Automation )	1025	47.23	11.49	5.42	51.3 mg/dL	0.75	0.34
3	<a href="#">CREATININE II</a>	Jaffes Kinetic- Alkaline Picrate II	Any Analyser (Automation / Semi Automation )	1158	1.04	15.24	0.16	1.1 mg/dL	0.38	0.01
4	<a href="#">T.BILIRUBIN II</a>	Diazonium Salt ( Colorimetric ) / Jendrassik II	Any Analyser (Automation / Semi Automation )	1337	4.71	15.27	0.72	4.6 mg/dL	-0.15	0.04
5	<a href="#">T-PROTEIN II</a>	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation )	1248	5.94	9.76	0.58	6.1 g/dL	0.28	0.03
6	<a href="#">ALBUMIN II</a>	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation )	1229	3.50	9.29	0.32	3.4 g/dL	-0.31	0.02
7	<a href="#">URIC ACID II</a>	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation )	1129	4.67	17.89	0.84	6.1 mg/dL	1.7	0.05
8	<a href="#">CHOLESTROL II</a>	CHOD-PAP II	Any Analyser (Automation / Semi Automation )	1590	128.46	11.58	14.88	133.3 mg/dL	0.33	0.75

9	<b>TRIGLYCERIDE II</b>	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation )	1346	160.03	14.05	22.49	198.1 mg/dL	1.69	1.23
<b>SDI Range</b>		<b>Interpretation</b>								
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

**Links**

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>