

CENTRIFUGE Calibration / Verification Record Sheet

Location: ICTC-CHC-NANDIKOTKUR-KURNOOL

Range: Up to 3500 RPM

Equipment ID: APSACS/ICTC/CHC/NDK/CEF/01

SL.NO: 20060142

Calibration Procedure Reference: In-house verification of centrifuge using certified standard calibrated by Tachometer

Date of Calibration:

30.05.2023

Due date of Calibration:

29.05.2024

Tachometer ID: KMC/MB/HIVTL/SRL/TM/01

Calibration date:
19.04.2023

Due date of Calibration: 18.04.2024

Verification Done by: B.BHASKAR, Technical Officer, SRL, KMC,KURNOOL

S.NO	STANDARD READING(RPM)	UUC READING (RPM)	ERROR (RPM)
1.	499.4	500	0.6
2.	998.7	1000	1.3
3.	1997.6	2000	2.4
4	2996.7	3000	3.3
5.	3496.1	3500	3.9

Accept/ Reject: Accepted since average difference is within +/- 100RPM

Final Interpretation: Centrifuge is in good working condition

Action Taken: - NA

Remark: - NA


B. Bhaskar
 TECHNICAL OFFICER