



CALIBRATION REPORT

Customer Name: Central Research Institute of Unani Medicine

Model: Fully Automated Bio Chemistry Analyzer EM-200

Serial No.: v200880

Calibration done date: 20/04/2023

Next Calibration due Date: On or Before 20/04/2024

Lab In charge: Dr. Mohd Naime

*This is to certify that above mentioned product has been verified of calibration
By
Checking Lamp Voltage 11.8V*

Checking gain of Photometer for all wavelength values obtained are as under:

Filter	340	405	505	546	578	600	660	700
Gain	476	470	470	497	484	499	500	479

Checked calibration of probe on all locations.

*Checked temperature at Reagent tray at 8.0 degree (Range 4 to 11 degree) and
reaction tray 37.2 degree (Range 36.5 to 37.5 Degree)*

*Followed by calibration of all the three types of chemistries (picked up CHOL,
CREATININE and SGPT) and QC run with satisfactory values of QC*

Calibration at site performed by Mr. Sachin Singh

Authorized Signatory
(Mr. Sachin Singh)
Sr. Area Service Engineer
For Transasia Bio-Medicals Ltd.