

CENTRIFUGE Calibration / Verification Record Sheet

Location: ICTC-AH-ADONI-KURNOOL		Range : Up to 3500RPM	
Equipment ID: APSACS/ICTC/CHC/ALU/CEF/01		SL.NO :19070228	
Calibration Procedure Reference: In-house verification of centrifuge using certified standard calibrated by Tachometer			
Date of Calibration :	05.09.2023	Due date of Calibration:	04.09.2024
Tachometer ID : KMC/MB/HIVTL/SRL/TM/01		Calibration date : 19.04.2023	Due date of Calibration : 18.04.2024

Verification Done by: **B.BHASKAR**, Technical Officer, SRL, KMC,KURNOOL


S.NO	STANDARD READING(RPM)	UUC READING (RPM)	ERROR (RPM)
1.	502	500	-2.0
2.	996.5	1000	4.5
3.	1997.1	2000	3.9
4.	2991.9	3000	9.1
5.	3496.8	3500	3.2

Accept/ Reject: Accepted since average difference is within +/- 100RPM

Final Interpretation: Centrifuge is in good working condition

Action Taken: - NA

Remark: - NA


 TECHNICAL OFFICER