

**Timer Calibration / Verification Record Sheet**

Location: ICTC-AH-ADONI

Equipment ID: APSACS/ICTC/AH/ADN/TIMER/01

Calibration Procedure Reference: In-house verification of timer using certified standard calibrated Timer

Date of Calibration :

05.09.2023

Due date of  
Calibration:

04.09.2024

Certified timer ID: APSACS/KMC/SRL/TIM/01 Calibration date : 05/04/2023

Due date of Calibration : 04/04/2024

Verification Done by: B.BHASKAR, Technical Officer, SRL, KMC, KURNOOL


Sr.No.	Timer Setting (Sec)	Certified Time (Sec)	Time Difference (Sec)	Timer Setting (Sec)	Certified Time (Sec)	Time Difference (Sec)	Timer Setting (Sec)	Certified Time (Sec)	Time Difference (Sec)
1	300	300	0	600	599	-1	900	900	0
2	300	299	-1	600	601	1	900	900	0
3	300	300	0	600	601	1	900	899	-1
4									
5									
Mean	300	299.6	-0.4	600	600.3	0.3	900	899.9	-0.1
SD	-	0.58	-	-	1.15	-	-	0.58	-
%CV	-	0.19	-	-	0.19	-	-	0.064	-
% Accuracy	-	-0.13	-	-	0.05	-	-	-0.011	-

Accept/ Reject: Accepted since average difference is within +/- 60 Second

Final Interpretation: Timer is in good working condition

Action Taken: - NA

Remark: - NA

  
TECHNICAL OFFICER