

CENTRIFUGE Calibration / Verification Record Sheet

Location: ICTC-GGH-MACHILIPATNAM

Range : Up to 3500RPM

Equipment ID: APSACS/ICTC/GGH/MTM/CEF/01

SL.NO : 02

Calibration Procedure Reference: In-house verification of centrifuge using certified standard calibrated by Tachometer

Date of Calibration : 22.05.2023 Due date of Calibration : 21.05.2024

Certified Reference ID : APSACS/ICTC/GGH/MTM/CEF/01 Technician ID : KNICAMB/HIVTL/SRL/TM/01 Calibration date : 19.04.2023 Due date of Calibration : 18.04.2024

Verification Done by: D.SUBBA RAO, Technical Officer, SRL, GGH,VIJAYAWADA

S.NO	STANDARD READING(RPM)	UUC READING (RPM)	ERROR (RPM)
1.	499.7	500	0.3
2.	999.2	1000	0.8
3.	1998.7	2000	1.3
4.	2997.5	3000	2.5
5.	3496.8	3500	3.2

Accept/ Reject: Accepted since average difference is within +/- 100RPM

Final Interpretation: Centrifuge is in good working condition

Action Taken: - NA

Remark: - NA

D.S.R.
22/05/23
TECHNICAL OFFICER
HIV TESTING LABORATORY
DEPARTMENT OF MICROBIOLOGY
SUDDHARTHA MEDICAL COLLEGE
VIJAYAWADA, A.P.-520 008.

Medical Officer
Integrated Counseling
and Testing Center (ICTC)
Government General Hospital
MACHILIPATNAM