



ISO 9001 : 2015  
RQ91/3144

# S.M. ENTERPRISES

(Calibration and Testing of Electrical, Electronic & Precision Instruments)

27<sup>th</sup>  
Years of Service

Office/Cal. Lab : Plot No. 12, "Ganesh Niwas", Sanjeevaiah Nagar, Co-operative Society, Besides Tadbund Hanuman Temple, Sikh Village, Secunderabad - 500 009, T.S, INDIA ☎: (040) 27840697, 48541008  
Cell: 9848020798, 9948291008, 9948292008, 9848293008, 9948294008 Email: msudheer\_98@yahoo.com, msudheer\_98@smentp.org, SME\_003@YAHOO.IN, url : www.smentp.org

## CALIBRATION CERTIFICATE

CALIBRATION ON 15/05/2023	CALIBRATION DUE 14/05/2024
1. Calibration Report Ref. No. & Date :	SME/NP/SMC/TIM/019/2023;Dt:15/05/2023
2. Requested by : (Name and address of the organization)	To, Prof & HOD( Department Of Microbiology). Siddhartha Medical College,Vijayawada.
3. Instrument Name:	Digital Timer
4. Instrument Make/Model:	ACETEQ / DT-1045
5. Instrument ID. No:	SMC/TIM/019
6. Range :	0 to 99 min
7. Accuracy :	+/- 1 % FSD
8. Environmental Condition	Temp/RH : 28 +/- 2 Deg C; R.H: 65 +/- 5%
9. Test parameter & condition :	Time V/S Time - (Comparison Method)
10. Standard Instruments used for Calibration:	Time Interval Meter
11. Traceability Certificate Ref. No.	C-220622-5-1 ,Date: 29/06/2022 , Due Dt: 29/06/2023

### TEST RESULTS


Sl No	Standard Reading Of Time	Observed Reading Of Time	Error Value(+/-)
	Hr:min:sec:ms	min:sec	Hr:min:sec:ms
1	00:01:00:29	01:00	00:00:00:29
2	00:10:00:35	10:00	00:00:00:35
3	00:30:00:19	30:00	00:00:00:19
4	01:50:00:44	90:00	00:00:00:44

PLEASE NOTE: THE INSTRUMENTS ERROR FOUND TO BE WITHIN THE SPECIFIED ACCURACY LIMIT AND THIS CERTIFICATE VALID FOR PERIOD OF ONE YEAR FROM THE DATE OF ISSUE.

CALIBRATED BY

S . M. Enterprises

ENGINEER INCHARGE



GSTIN : 36AFVPS4924F1ZW Service Tax No. : AFVPS4924FST001 PAN No. : AFVPS4924F

NABL APPROVED LAB (CC-2471) / Approved Laboratory of ISO 9001-2015

TEST EQUIPT. / STANDARDS USED ARE TRACEABLE TO NATIONAL STANDARDS