



**PROFICIENCY TESTING REPORT**  
**ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME**  
 NABL accredited program as per ISO/IEC 17043:2010 standard  
 Organized By Department of Hematology, AIIMS, New Delhi-110029



*Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens*

EQAP CODE No. : 3297

Distribution No.: 160

Month/Year: March/2023

Instrument ID: BC30(UD94000739)

Name & Contact No. of PT Co-ordinator: Dr. Seema Tyagi (Prof.), Hematology, AIIMS, Delhi,  
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue &amp; status of the report: 11-05-2023[Final].

### CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 <sup>3</sup> /µl	1	5.6	5.5	11.1	13.6	0.109	-1.37	0.1	0.14	0.018	-0.21
RBC x10 <sup>6</sup> /µl	1	4.13	4.1	8.23	8.31	0.014	-0.32	0.03	0.05	0.051	-0.39
Hb g/dl	1	11.9	11.9	23.8	25.25	0.054	-1.78	0	0.1	0.014	-0.67
HCT%	1	40.7	40.3	81	80.1	0.318	0.16	0.4	0.5	0.047	-0.19
MCV-fl	1	98.4	98.4	196.8	193.95	0.700	0.26	0	0.4	0.045	-0.90
MCH-Pg	1	29.1	28.8	57.9	60.85	0.123	-1.47	0.3	0.3	0.035	0.00
MCHC-g/dl	1	29.6	29.3	58.9	62.9	0.249	-0.93	0.3	0.3	0.029	0.00
Plt. x10 <sup>3</sup> /µl	1	195	193	388	410.5	2.109	-0.58	2	6	0.656	-0.54
Retic %	2			0							

### P.S . Assesment

YOUR REPORT			CONSENSUS REPORT		
DLC%	3	Nrbcs= , Poly= L=, E=, Mono/Promono= , B1= P.M.=, Mye=, Meta=, Other=	Poly: 25-45, nRBC/Lymph/Mono/Eo/Pro/Blast: 0-5, Myelo: 20-40, Meta: 10-20, Baso: 0-4		
RBC Morphology	3		Predominantly: Normocytic Normochromic, Anisocytosis. Moderate: Macrocytic. Mild: Microcytic		
Diagnosis	3		Chronic Myeloid Leukemia-Chronic Phase [ CML-CP]		

**COMBINED DATA VALUES OF TOTAL PARTICIPANTS**

Test parameters	S.No.	Total participants covered in the current dist. 160	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
<b>WBC x10<sup>3</sup>/µl</b>	1	114	108	87.96	91.67	5.56	1.85	6.48	6.48
<b>RBC x10<sup>6</sup>/µl</b>	1	114	114	88.6	81.58	3.51	4.39	7.89	14.03
<b>Hb g/dl</b>	1	114	114	85.96	53.51	4.39	28.95	9.65	17.54
<b>HCT%</b>	1	114	108	94.44	87.04	4.63	3.7	0.93	9.26
<b>MCV-fl</b>	1	114	108	93.52	95.37	3.7		2.78	4.63
<b>MCH-Pg</b>	1	114	108	89.81	88.89	7.41	4.63	2.78	6.48
<b>MCHC-g/dl</b>	1	114	108	95.37	88.89	3.7	5.56	0.93	5.55
<b>Plt. x10<sup>3</sup>/µl</b>	1	114	108	89.81	92.59	5.56	2.78	4.63	4.63
<b>ReticCount%</b>	2	114	94	89.36	84.04	5.32	4.26	5.32	11.70
<b>PS Assessment</b>	3	114	108	Satisfactory :82.8%, Borderline Sat. :27.2%, Unsatisfactory :0%					

**\*Comments:**

**1). Among Lab (EQA) : Results acceptable.**

**2). Within Lab (IQA) : Precision acceptable.**

**Note-1: EQA** (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

**IQA** ( Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

**Note-2:** Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

**Note-3:** Z score 0 to  $\pm 2$ : Acceptable, Z score  $\pm 2$  to  $\pm 3$  :Warning Signal, Z score  $> \pm 3$  : Unacceptable [As per ISO/IEC 13528:2015 standard]

**Note-4:** Z score value between " $0$  to  $\pm 2$ " are texted in green colour. Z score value between " $\pm 2$  to  $\pm 3$ " are texted in orange colour. Z score value  $> \pm 3$  are texted in red colour.

**Note-5:** Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value ( $0.3*SDPA$ ). To pass the stability test, average difference in measurement values of first and last day sample ( $\bar{x}-\bar{y}$ ) should be smaller than the check value ( $0.3*SDPA$ ).

**Note-6:** ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

**Note-7:** Participants are free to use methods/analyzer of their own choice.

**Note-8:** Proficiency testing (PT ) samples are sent quarterly to each participant.

**Note-9:** All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website [www.ishtmaiimseqap.com](http://www.ishtmaiimseqap.com).

**Note 10:** Reports are kept confidential.

Report authorized by,



Dr. Seema Tyagi (Prof.)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----