



UNIQUE CALIBRATION & CONSULTANT SERVICES

AN ISO 9001:2015 CERTIFIED CALIBRATION LABORATORY



Calibration Certificate

Page No :- 1 of 1

Format No.: UCCS/QF7.8-07

Certificate No.	UCCS/21/A/167-011	Calibration Certificate of Micro pipette
Issue Date	11.03.2023	
Service Request No.	UCCS/21/H/168	
Service Request Date	11.03.2023	
Receipt Date	11.03.2023	
Calibration Date	11.03.2023	
Calibration Due Date (As per Customer request)	10.03.2024	

Name & Address of Customer	Identification on UUC (Unit Under Calibration)	
M/s. XTRA CARE DIAGNOSTICS CHEMBUR MUMBAI -74	Range/ Size	10 to 100 ul
	Make/Model	Perfect
	Instrument I.D. No.	Xtra/REF-T/ML/13
	Instrument Sr. No.	
	Least Count	0.1 ul
	Location/Dept.	MAIN LAB
	Visual Inspection	Ok

Calibration Procedure	UCCS/TH/CP- 04	Condition of Instrument	OK
Reference Standard	DKD R5-1	Calibration Performed at	Lab
Environmental Condition	Temperature	30 ± 5 °C	Relative Humidity 30 to 80 % RH

Standard Equipment Used for Calibration					
Instrument Name	Make	I.D No./Sr.No.	Calibrated By	Certificate No.	Due Date
Tycon Engineering	TIPL	UCCL/ET/PC-01	Acumen Measurement Delhi	AMC/00001537	6/30/2024

Serial No	Nominal Value In (kg/cm ²)	Observed Value In kg/cm ²	Error in °C Mean	Uncertainty of Measurement ± °C
1	20.0	19.9	0.1	± 0.46 °C
2	40.0	39.8	0.2	
3	80.0	79.8	0.2	
4	100.0	99.9	0.1	

The reported uncertainty stated is the expanded uncertainty of measurement obtained by multiplying the standard uncertainty by the coverage factor k=2 correspond to confidence level of 95 %

- Note :-
1. This Certificate refers only to the particular items submitted for calibration.
 2. This Certificate shall not be reproduced, except in full, without the written permission of Managing director Calibration System, Rudrapur.
 3. Result Reported are Valid at the of and under the stated conditions of measurement.
 4. Laboratory Standards are traceable to National Standards.
 5. Calibration Certificate issue for weight & Measure parameters i.e. Mass, Balance, Volumetric equipment,



Measuring Scale/Tapes etc. for scientific purpose only and should not be used for Trade / Commercial use.

End of Certificate

Amrit
Calibrated by
Amrit Kumar
Calibration Engineer



Reviewed/Approved By
Vikash Kumar
Technical Manager

Working : W.Z. -1390/Z-2, 1st Floor, Nangal Raya Extn., Pankha Road, New Delhi-110046
Regd. Office : R.Z.-59A, Indra Park, Gali No.25, Palam Colony, New Delhi-110045