



# UNIQUE CALIBRATION & CONSULTANT SERVICES

AN ISO 9001:2015 CERTIFIED CALIBRATION LABORATORY



## Calibration Certificate

Formal No.: UCCS/OF/17.8-05

Page No :- 1 of 1

Certificate No. Issue Date Service Request No. Service Request Date Receipt Date Calibration Date Calibration Due Date (As per Customer request)	UCCS/21/A/167-009 11.03.2023 UCCS/21/H/166 11.03.2023 11.03.2023 11.03.2023 10.03.2024	<b>Calibration Certificate of</b>  <b>Micro pipette</b>
--	--	---

Name & Address of Customer	Identification on UUC (Unit Under Calibration)	
M/s. XTRA CARE DIAGNOSTICS CHEMBUR MUMBAI -74	Range/ Size Make/Model Instrument I.D. No. Instrument Sr. No. Least Count Location/Dept. Visual Inspection	100 to 1000 ul Perfect Xtra/REF-T/ML/11 0.1 ul MAIN LAB Ox

Calibration Procedure	UCCS/TH/CP- 04	Condition of Instrument	OK
Reference Standard	DKD R5-1	Calibration Performed at	Lab
Environmental Condition	Temperature	30 ± 5 °C	Relative Humidity 30 to 80 % RH

Standard Equipment Used for Calibration					
Instrument Name	Make	I.D No./Sr.No.	Calibrated By	Certificate No.	Due Date
Tycon Engineering	TIPL	UCCL/ET/PC-01	Acumen Measurement Delhi	AMC/00001537	6/30/2024

Serial No	Nominal Value In (kg/cm2)	Observed Value In kg/cm2	Error in °C Mean	Uncertainty of Measurement ± °C
1	200.0	199.9	0.1	± 0.46 °C
2	500.0	499.8	0.2	
3	1000.0	999.7	0.4	

The reported uncertainty stated is the expanded uncertainty of measurement obtained by multiplying the standard uncertainty by the coverage factor k=2 correspond to confidence level of 95 %

- Note :-
1. This Certificate refers only to the particular items submitted for calibration.
  2. This Certificate shall not be reproduced except in full without the written permission of Managing director Calibration System, Rudrapur.
  3. Result Reported are Valid at the of and under the stated conditions of measurement.
  4. Laboratory Standards are traceable to National Standards.
  5. Calibration Certificate issue for weight & Measure parameters i.e. Mass, Balance, Volumetric equipment,



Measuring Scale/Tapes etc. for scientific purpose only and should not be used for Trade / Commerical use.

\*\*\*End of Certificate\*\*\*

Authorized By  
Amit Kumar  
Calibration Engineer

Authorized Signatory  
Approved By  
Vikash Kumar  
Technical Manager

Working : W.Z. -1390/Z-2, 1st Floor, Nangal Raya Estn., Pankha Road, New Delhi-110046  
Regd. Office : R.Z.-68A, Indra Park, Gali No.25, Palam Colony, New Delhi-110045