

**Community Health Centre, Vuyyuru, Krishna District, A.P.**

**Legal Entity of Integrated Counselling and Testing Center (ICTC)**

The ICTC is established on March 2005 in collaboration with National AIDS Control Organization and AP State AIDS Control Society in Community Health Centre, Vuyyuru. The ICTC is rendering testing and counselling services to its clients from the premises of Community Health Centre, Vuyyuru.

We have here authorized the legal entity of ICTC which located in CHC, Vuyyuru, functional, and rendering services in the premises of the Community Health Centre, Vuyyuru, under the supervision of the Medical Superintendent.

*Srinivas Reddy*  
Medical Officer  
VCTCT, PPTCT  
P. R. M. Govt. Hospital  
Vuyyuru

Medical Superintendent,  
Community Health Centre

Vuyyuru

108773

ANDHRA PRADESH



**Dr. NTR UNIVERSITY OF HEALTH SCIENCES**  
**FACULTY OF MODERN MEDICINE**

**This is to certify that**

**TEJASWI PONNAM**

**has been awarded the Degree of**

**Bachelor of Medicine**

**and**

**Bachelor of Surgery**

*He/She having been declared to have qualified in the Examination prescribed therefore, of this University held in MARCH, 2012*


*and on completion of one year Compulsory Rotatory Internship. He/She has been placed in FIRST Division.*

*Given under the Seal of the University*

Regd. No. 0763124

VIJAYAWADA

Dated 5-5-2013





Registrar

RECEIVED BY THE DIRECTOR OF THE BUREAU OF THE CENSUS  
WASHINGTON, D. C. 20543



100-100000-100000

100-100000-100000

100-100000-100000

100-100000-100000

100-100000-100000

100-100000-100000

100-100000-100000

100-100000-100000

100-100000-100000

100-100000-100000

100-100000-100000

100-100000-100000

100-100000-100000



# S.M. ENTERPRISES 27<sup>th</sup>

ISO 9001:2015  
RSM/3144

(Calibration and Testing of Electrical Electronic & Precision Instruments)

Waste of Service

Head Office: 10/10/2, Ganesh Wasti, Sanyalpur Nagar, Postcode: Sahasra, Sector: Sahasra, Dist: Sahasra, Pin: 731104, (WB) INDIA  
Tel: 9830781889, 9830781889, 9830781889 Email: smenterprises@gmail.com Website: www.smenterprises.com

## CALIBRATION CERTIFICATE

CALIBRATION ON  
15/05/2023

CALIBRATION DUE  
14/05/2024

1. Calibration Report Ref. No. & Date :

SME/NT/SMC/DTHM/011/2023/DT/15.05.2023

2. Requested by :

Ta. Prof & HOD, Department of Microbiology 1

3. Instrument Name

Siddhartha Medical College, Vijaywada

4. Make/Model:

Digital Thermo Hygrometer  
HTC-2

5. Instrument ID No:

SMC/DTHM/011

6. Range :

-100°C to 50.0 °C & 10% to 99 % RH

7. Accuracy :

± 1 °C & 3% RH

8. Environmental Conditions :

Temp: RH: 27 °C, 20% C, RH: 60 % - 5 %

9. Test Parameter & condition :

Temperature & RH V/S Temperature& RH comparison method

10. Standard Instruments used for Calibration:

Digital Thermo Hygrometer

11. Traceability Certificate Ref. No.

ISC/22-23/ISS-7-1, Date: 06/09/2022, Due Date: 06/09/2023

### TEST RESULTS

| Sl No | Standard Reading Of |                   | Observed Reading Of Under             |         | Error Value |         |
|-------|---------------------|-------------------|---------------------------------------|---------|-------------|---------|
|       | Temperature & RH    | Temperature in °C | Calibration Hygrometer In Deg C & %RH | RH in % | Temp in °C  | RH in % |
| 1     | Temperature in °C   | -10.0             | Temperature in °C                     | -10.0   | Temp in °C  | 0.0     |
| 2     | RH in %             | 10                | RH in %                               | 12      | RH in %     | 2       |
| 3     | Temperature in °C   | 0.0               | Temperature in °C                     | 0.1     | Temp in °C  | 0.1     |
| 4     | RH in %             | 10.0              | Temperature in °C                     | 10.2    | Temp in °C  | 0.2     |
| 5     | Temperature in °C   | 20.0              | Temperature in °C                     | 19.9    | Temp in °C  | 0.1     |
| 6     | RH in %             | 50                | Temperature in °C                     | 29.8    | Temp in °C  | 0.2     |
| 7     | Temperature in °C   | 30.0              | Temperature in °C                     | 29.8    | Temp in °C  | 0.2     |
| 8     | RH in %             | 40.0              | Temperature in °C                     | 29.8    | Temp in °C  | 0.2     |
| 9     | Temperature in °C   | 50.0              | Temperature in °C                     | 49.8    | Temp in °C  | 0.2     |
| 10    | RH in %             | 90                | RH in %                               | 88      | RH in %     | 2       |


PLEASE NOTE: THE INSTRUMENTS ERROR NOT FOUND TO BE WITHIN THE SPECIFIED ACCURACY LIMIT AND THIS CERTIFICATE VALID FOR PERIOD OF ONE YEAR FROM THE DATE OF ISSUE.

CALIBRATED BY  
S.M. ENTERPRISES



GSTIN : 36AFVPS4924F1ZW Service Tax No. : AFVPS4924FST001 PAN No. : AFVPS4924F  
NABL APPROVED LAB (CC-2471) / Approved Laboratory of ISO 9001:2015  
TEST EQUIPT. / STANDARDS USED ARE TRACEABLE TO NATIONAL STANDARDS

| CENTRIFUGE Calibration / Verification Record Sheet   |                       |  |                          |                               |
|--|-----------------------|--|--------------------------|-------------------------------|
| Location: ICTC-CHC-VUYYURU-KRISHNA DIST,A,P  |                       |  | Range : Up to 3500RPM    |                               |
| Equipment ID: APSACS/ICTC/CHC/VYR/CEF/01   |                       |  | SL.NO :                  |                               |
| Calibration Procedure Reference: In-house verification of centrifuge using certified standard calibrated by Tachometer |                       |  |                          |                               |
| Date of Calibration :  | 22.08.2023            |  | Due date of Calibration: | 21.08.2024                    |
| Certified centrifuge ID : APSACS/SMC/SRL/CEF/01  |                       | Tachometer ID : SMC/MB/HIVTL/SRL/TM/01 |                          | Calibration date : 19.08.2023 |
| Verification Done by: D.SUBBARAO, Technical Officer, SRL, SMC,VIJAYAWADA.  |                       |  |                          |                               |
| S.NO   | STANDARD READING(RPM) | UUC READING (RPM)                      | ERROR (RPM)              |                               |
| 1.   | 999.5                 | 1000                                   | 0.5                      |                               |
| 2.   | 1999.1                | 2000                                   | 0.9                      |                               |
| 3.   | 3998.3                | 4000                                   | 1.7                      |                               |
| 4  | 4997.5                | 5000                                   | 2.5                      |                               |
| 5.   | 5996.2                | 6000                                   | 3.8                      |                               |
| Accept/ Reject: Accepted since average difference is within +/- 100RPM   |                       |  |                          |                               |
| Final Interpretation: Centrifuge is in good working condition  |                       |  |                          |                               |
| Action Taken: - NA   |                       |  |                          |                               |
| Remark: - NA   |                       |  |                          |                               |
|  |                       |  |                          |                               |
|  |                       |  |                          |                               |

  
**TECHNICAL OFFICER**  
 HIV TESTING LABORATORY  
 DEPARTMENT OF MICROBIOLOGY  
 SIDDHARTHA MEDICAL COLLEGE  
 VIJAYAWADA, A.P.-520 008.