

CENTRIFUGE Calibration / Verification Record Sheet

Location: PPTCT-GGH-KADAPA

Range : Up to 3500RPM

Equipment ID: APSACS/PPTCT/GGH/KDP/CEFF/01

SL.NO : 02

Calibration Procedure Reference: In-house verification of centrifuge using certified standard calibrated by Tachometer

Date of Calibration :	26.08.2023	Due date of Calibration:	25.08.2024
------------------------------	------------	---------------------------------	------------

Certified centrifuge ID : GMC-ATP/HIVTL/SRL/CEFF/01	Tachometer ID : GMC-ATP/HIVTL/SRL/TM/01	Calibration date :	19.05.2023	Due date of Calibration :	18.05.2024
--	--	---------------------------	------------	----------------------------------	------------

Verification Done by: P. ANURADHA, Technical Officer, SRL, GMC- ANANTHAPURAMU

S.NO	STANDARD READING(RPM)	UUC READING (RPM)	ERROR (RPM)
1.	499.7	500	0.3
2.	999.2	1000	0.8
3.	1998.7	2000	1.3
4	2998.5	3000	1.5
5.	3497.8	3500	3.0

Accept/ Reject: Accepted since average difference is within +/- 100RPM

Final Interpretation: Centrifuge is in good working condition

Action Taken: - NA

Remark: - NA

P. Anuradha
SRL-7.0
Dept. of Microbiology
GMC - Ananthapuram