

**CENTRIFUGE Calibration / Verification Record Sheet**

**Location:** ICTC-GGH-KADAPA

**Range :** Up to 3500RPM

**Equipment ID:** APSACS/ICTC/GGH/KDP/CEF/01

**SL.NO :** 02

**Calibration Procedure Reference:** In-house verification of centrifuge using certified standard calibrated by Tachometer

**Date of Calibration :** 26.08.2023 **Due date of Calibration:** 25.08.2024

**Certified centrifuge ID :** GMC-ATP/HIVTL/SRL/CEF/01 **Due date of Calibration :** 18.05.2024

**Verification Done by:** P. ANURADHA, Technical Officer, SRL, GMC- ANANTHAPURAMU **Calibration date :** 19.05.2023

S.NO	STANDARD READING(RPM)	UUC READING (RPM)	ERROR (RPM)
1.	499.7	500	0.3
2.	999.2	1000	0.8
3.	1998.7	2000	1.3
4	2999.5	3000	0.5
5.	3496.8	3500	3.2

**Accept/ Reject:** Accepted since average difference is within +/- 100RPM

**Final Interpretation:** Centrifuge is in good working condition

**Action Taken:** - NA

**Remark:** - NA

*P. Anuradha*  
*SRL - T.O*  
*Dept. of Microbiology*  
*GMC - Ananthapuram.*