S.M. ENTERPRISES
Certificate Traceable to NABL
CALIBRATION CARD
Cal. Dt. 25/05/2023 24/05/2024
Inst. Ref. No. SVMC/TIM/016
Instrument within/out of Accuracy Limit

ENTERPRISES

27th Years of Service

d Testing of Electrical, Electronic & Precision Instruments)

ar, Co-operative Society, Besides Tadbund Hanuman Temple, Sikh Village, Secunderabad - 500 009, T.S. INDIA ☎: (040) 27840697, 48541008
0008, 9948294008 Email: msudheer_98@yahoo.com, msudheer_98@smentp.org, SME_003@YAHOO.IN, url : www.smentp.org

CALIBRATION CERTIFICATE

CALIBRATION ON 25/05/2023	CALIBRATION DUE 24/05/2024
1. Calibration Report Ref. No. & Date :	SME/NP/SVMC/TIM/016/2023;Dt:25/05/2023
2. Requested by :	**To, Prof & HOD (State reference Lab)**
(Name and address of the organization)	Department Of Microbiology SV Medical College Tirupathi AP
3. Instrument Name:	Digital Timer
4. Instrument Make/Model:	Aceteq / DT-1045
5. Instrument ID. No:	SVMC/TIM/016
6. Range :	0 to 99 min
7. Accuracy :	+/- 1 % FSD
8. Environmental Condition	Temp/RH : 28 +/- 2 Deg C; R.H: 65 +/- 5%
9. Test parameter & condition :	Time V/S Time (Comparision Method)
10. Standard Instruments used for Calibration:	Time Interval Meter
11. Traceability Certificate Ref. No.	C-220622-5-1 ,Date: 29/06/2022 , Due Dt: 29/06/2023

TEST RESULTS

Sl No	Standard Reading Of Time	Observed Reading Of Time	Error Value(+/-)
	Hr:min:sec:ms	min:sec	Hr:min:sec:ms
1	00:01:00:32	01:00	00:00:00:32
2	00:10:00:38	10:00	00:00:00:38
3	00:30:00:39	30:00	00:00:00:39
4	01:50:00:46	90:00	00:00:00:46

PLEASE NOTE: THE INSTRUMENTS ERROR FOUND TO BE WITHIN THE SPECIFIED ACCURACY LIMIT AND THIS CERTIFICATE VALID FOR PERIOD OF ONE YEAR FROM THE DATE OF ISSUE.

CALIBRATED BY
S.M. Enterprises
ENGINEER IN CHARGE

S.M. ENTERPRISES An ISO 9001-2015 Co. Incharge Reg. No. RQ9 1144 SECUNDERABAD