

Testing of Electrical, Electronic & Precision Instruments)

Co-operative Society, Besides Tadband Hanuman Temple, Sikh Village, Secunderabad- 500 009, T.S, INDIA. e: (P)01 27840097, 48541008
 93006, 9948294006 Email: msudheer.98@yahoo.com, msudheer.98@smerp.org, SME_003@YAHOO.IN, url: www.smerp.org

CALIBRATION CERTIFICATE

CALIBRATION ON 25/05/2023	CALIBRATION DUE 24/05/2024
1. Calibration Report Ref. No. & Date :	SME/NP/SVMC/DTHM/CAL-016/2023;Dt:25/05/2023
2. Requested by :	To, Prof & HOD (State reference Lab) Department Of Microbiology SV Medical College Tirupathi AP
3. Instrument Name	Digital Thermo Hygrometer
4. Make/Model:	HTC-2
5. Instrument ID No:	SVMC/DTHM/016
6. Range :	-10.0°C to 50.0 °C & 10% to 99 % RH
7. Accuracy :	+/- 1° C & 5%RH
8. Environmental Conditions :	Temp/ RH : 27 +/- 2Deg C; R.H:- 60 +/- 5 %
9. Test parameter & condition :	Temperature & RH V/S Temperature& RH compresion method
10. Standard Instruments used for Calibration:	Digital Thermo Hygrometer
11. Traceability Certificate Ref. No.	TSC/22-23/8627-1, Date:06/09/2022, Due Date:06/09/2023.

TEST RESULTS

Sl No	Standard Reading Of Temperature & RH		Observed Reading Of Under Calibration Hygrometer In Deg C & %RH		Error Value +/-	
	Temperature in °C	RH in %	Temperature in °C	RH in %	Temp in °C	RH in %
1	-10.0	10	-10.0	11	0.0	1
2	0.0	30	0.1	32	0.1	2
3	10.0	40	10.1	38	0.1	2
4	20.0	50	19.9	52	0.1	2
5	30.0	60	29.8	62	0.2	2
6	40.0	75	39.8	74	0.2	1
7	50.0	90	49.9	87	0.1	3

PLEASE NOTE: THE INSTRUMENTS ERROR NOT FOUND TO BE WITHIN THE SPECIFIED ACCURACY LIMIT AND THIS CERTIFICATE VALID FOR PERIOD OF ONE YEAR FROM THE DATE OF ISSUE.

CALIBRATED BY
 S.M. ENTERPRISES

