

CENTRIFUGE Calibration / Verification Record Sheet

Location: ICTC-CHC-JEGGALAHPETTA-NTR-DIST.A.P

Range : Up to 3500RPM

Equipment ID: APSACS/ICTC/CHC/JPT/CEF/01

SL.NO :01

Calibration Procedure Reference: In-house verification of centrifuge using certified standard calibrated by Tachometer

Date of Calibration :

11.05.2023

Tachometer ID : SMC/MB/HIV/TL/SRL/TM/01

Due date of Calibration:

10.05.2024

Verification Done by:

D.SUBBARAO, Technical Officer, SRL, SMC,VIJAYAWADA.

Calibration date :

Due date of Calibration : 18.04.2024

S.NO	STANDARD READING(RPM)	UUC READING (RPM)	ERROR (RPM)
1.	499.3	500	0.7
2.	998.7	1000	1.3
3.	1997.6	2000	2.4
4	2996.6	3000	3.4
5.	3496.2	3500	3.8

Accept/ Reject: Accepted since average difference is within +/- 100RPM

Final Interpretation: Centrifuge is in good working condition

Action Taken: - NA

Remark: - NA


11/05/23

TECHNICAL OFFICER

HIV TESTING LABORATORY
DEPARTMENT OF MICROBIOLOGY
SIDDHARTHA MEDICAL COLLEGE
VIJAYAWADA, A.P.-520 008.

TECHNICAL OFFICER

HIV TESTING LABORATORY
DEPARTMENT OF MICROBIOLOGY
SIDDHARTHA MEDICAL COLLEGE
VIJAYAWADA, A.P.-520 008.