


CENTRIFUGE Calibration / Verification Record Sheet				
Location: ICTC-CHC-NANADIGAMA-NTR-DIST,A,P		Range : Up to 3500RPM		
Equipment ID: APSACS/ICTC/CHC/NDGM/CEF/01		SL.NO :01		
Calibration Procedure Reference: In-house verification of centrifuge using certified standard calibrated by Tachometer				
Date of Calibration :	21.06.2023	Due date of Calibration:	20.06.2024	
Tachometer ID : SMC/MB/HIVTL/SRL/TM/01		Calibration date :	19.04.2023	
Verification Done by: D.SUBBARAO, Technical Officer, SRL, SMC,VIJAYAWADA.				
S.NO	STANDARD READING(RPM)	UUC READING (RPM)	ERROR (RPM)	
1.	499.3	500	0.7	
2.	998.7	1000	1.3	
3.	1997.6	2000	2.4	
4.	2996.6	3000	3.4	
5.	3496.2	3500	3.8	
Accept/ Reject: Accepted since average difference is within +/- 100RPM				
Final Interpretation: Centrifuge is in good working condition				
Action Taken: - NA				
Remark: - NA				


TECHNICAL OFFICER
 HIV TESTING LABORATORY
 DEPARTMENT OF MICROBIOLOGY
 SIDHARTHAN MEDICAL COLLEGE
 VIJAYAWADA, A.P.-520 008.