

CENTRIFUGE Calibration / Verification Record Sheet

Location: ICTC-CHC-ALLAGADDA-KURNOOL

Range : Up to 6500RPM

Equipment ID: APSACS/ICTC/CHC/ALG/CEF/01

SL.NO : ZDHN23060

Calibration Procedure Reference: In-house verification of centrifuge using certified standard calibrated by Tachometer

Date of Calibration :

14.09.2023

Due date of Calibration:

13.09.2024

Certified centrifuge ID :

APSACS/KMC/SRL/CEF/01

Tachometer ID :

KMC/MB/HIVTL/SRL/TM/01

Calibration date :

19.04.2023

Due date of Calibration :

18.04.2024

Verification Done by: B.BHASKAR, Technical Officer, SRL, KMC, KURNOOL

S.NO	STANDARD READING(RPM)	UUC READING (RPM)	ERROR (RPM)
1.	499.7	500	0.3
2.	999.2	1000	0.8
3.	1998.7	2000	1.3
4	2997.5	3000	2.5
5.	3496.8	3500	3.2

Accept/ Reject: Accepted since average difference is within +/- 100RPM

Final Interpretation: Centrifuge is in good working condition

Action Taken: - NA

Remark: - NA


 TECHNICAL OFFICER