




CHRISTMAN MEDICAL COLLEGE
 VELLORE
In Service of the Nation Since 1900...



VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. Block PHC, Andiyappanur
 Lab No: 3038
 Month: July
 Year: 2023
 Constituent Group: Chemistry III

[Click on the analyte to view Graphical Data](#)

- [Details About Robust Analysis](#)
- [Detail About Monthly Summary](#)
- [Detail about SDI](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered : 20/07/2023

Date of Report Published : 07/08/2023

SI	Analyte	Principle Name	Analyzer Name	No of Participants	Mean	SD	CV	Participants	CV	SDI	CV
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	108 mg/dL	-0.22	0.28	
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	5.8 mg/dL	-0.09	0.02	
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	114 mg/dL	0.21	0.22	
SDI Range				Interpretation							
Within -1.00 to +1.00				Excellent.							
Within ±1.01 to ±2.00				Good.							
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.							
Beyond ±3.0				Unacceptable performance. Action Signal.							

[Links](#)