



CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23/08186	MEDICAL DEVICES	Page No:1 of 1
Issue Date	03-07-2023	
SRF No & Date	SRF/23/00246-0001 & 01-07-2023	
Receipt Date	01-07-2023	
Calibration Date	01-07-2023	
Calibration Due	30-06-2024	

Customer Name & Address
GOVERNMENT PRIMARY HEALTH CENTRE,
PUZHUTHIPATTY-630309, SINGAMPUNARI (TK), SIVAGANGAI DISTRICT.

Details of Device Under Calibration (DUC)			
Description	ELECTRICAL SAFETY (MICROSCOPE)	Make & Model	LABOMED & VISION 200
Range	MULTI	Sr. No	211032893
Resolution	MULTI	Identification No	
DUC Condition	SATISFACTORY	Location	LAB

Environmental Conditions & Calibration Procedure Details			
Environmental Details	Temperature: 25.6°C	Relative Humidity	54% RH
Calibration Procedure No	SBS/CP/MD/29	Calibration done at	ONSITE

Reference Standards Details				
S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	TSC/22-23/7400-3	10-08-2023

ELECTRICAL SAFETY

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resistance >20MΩ	Measured values in MΩ 96	Uncertainty in % (±) 13.92
2	Earth Leakage <5000µAfor B,BF,CF	Measured values in µA 210	Uncertainty in % (±) 7.1
3	Enclosure Leakage <500µAfor B,BF,CF	Measured values in µA 211	Uncertainty in % (±) 6.7


REMARKS

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
- Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

Authorised Signatory

C. Sivabalan
(Calibration Engineer)
C.SIVABALAN



C.SD
Technical Manager
(C.SHANMUGARAJ)

Chief Executive

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.