

## CALIBRATION CERTIFICATE

U/R Number: CC/15/190000837908

**Name and address of the customer**  
 The Medical Officer  
 Government Primary Health Centre  
 Karaiyur - pudukkottai (dist)

**Certificate no** Rml/01/31910/848/11  
**Date of calibration** 10/03/2023  
**Suggested calibration due** 10/03/2024  
**Customer po. reference & date**

**1 DESCRIPTION OF DUC**

**Name Of The Duc:** MICROSCOPE  
**Make** BINOCULAR/labomed  
**Model** LED/vision 2000  
**SLNo** 202220594  
**Range/Size** 230v

**Receiving Condition Of Duc** Good  
**Ref. Document** As Per Manufactures Specification  
**Environmental Condition** 25.3 Deg & 54.3% Rh  
**Uncertainty Of Measurement** As Per Report  
**Location** lab

**THE MASTERS ARE TRACEABLE TO NATIONAL/INTERNATIONAL STANDARDS AS PER ISO/IEC 17025.**

**2 DETAILS OF MASTER USED**

Standard Instrument Used	Make/ Model	Id No	Traceable To/Certificate No	Validity
Decade Megohm	Vaisashika	Rml-230	Vaiseshika,Ambala Cantt/Ved/Nabl/18-19(Et)/92	1-6-2023
Multi Product Calibrator	Fluke	Rml-262	Wavecal,Blr/Cs-18/01-01	1-6-2023
Multi Product Calibrator (Ac/Dc Current Coil)	Transmille	Rml-252-1	Wavecal,Blr/Cs-18/69-01	1-6-2023
Leakage Current Tester	Htc/CI-2054	Rml-254-1	Rml/01/31910/848	1-6-2023

**CALIBRATION FIELD:** Electro Technical (Voltage Calibration - Ac)

Sl.no	Range	Master readings v	Calibrated values v	Error observed v	Uncertainty v
1	0 v to 400v	230	231	1	0.055

**CURRENT CALIBRATION:** Ac Current @50Hz

Sl.no	Range	Master readings mA	Calibrated values mA	Error observed mA	Uncertainty mA
1	40Ma-100A	2	2.002	-0.002	0.045

**Resistance calibration**

Sl.no	Range	Master readings ohms	Calibrated values ohms	Error observed ohms	Uncertainty Ohms
1	40	39	39.01	0.001	0.007

**4.remarks**

- 4.1 uncertainty estimated for a confidence level of 95% at the k factor=2
  - 4.2 the reported results are valid only for the condition of the inst./gauge received at the time of and under the condition of the calibration
  - 4.3 DUC -device under calibration
- Calibration label /status. Calibration result pass & fail are indicated



 CALIBRATION ENGINEER	 CALIBRATION SEAL	 AUTHORIZED SIGN	This Report Refers Only To The Sample Submitted And May Not Be Reproduces In Full Or In Part, Without Written Permission From Roots Metrology & Testing Laboratory, Coimbatore, cell 9585840192
--	---	--	---