

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. Block PHC, Poovanthi

Details About Robust Analysis

Lab No 4547

Detail About Monthly Summary

Month August

Detail about SDI

Year 2023

Constituent Group Chemistry III

Click on the analyte to view Graphical Data

[All Analyser Result](#)[Print](#)

Date of Result Entered : 19/08/2023

Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	165 mg/dL	-1.18	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.7 mg/dL	-0.50	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	96 mg/dL	-3.84	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out