

Lab Name Govt. Block PHC, Poovanthi
 Lab No 4547
 Month July
 Year 2023
 Constituent Group Chemistry III

Details About Robust Analysis
 Detail About Monthly Summary
 Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 18/07/2023

Date of Report Published : 07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	103 mg/dL	-1.05	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	6.2 mg/dL	1.06	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	99 mg/dL	-3.00	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

