

CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23108775

MEDICAL DEVICES

Page No.1 of 1

Issue Date : 20-07-2023
 SRF No & Date : SRF/23/00264-0011 & 17-07-2023
 Receipt Date : 17-07-2023
 Calibration Date : 17-07-2023
 Calibration Due : 16-07-2024

Customer Name & Address:
 GOVERNMENT PRIMARY HEALTH CENTRE,
 POOVANTHI-630611,SIVAGANGAI DISTRICT.

Details of Device Under Calibration (DUC)

Description	MICROSCOPE (LIGHT)	Make & Model	LABMED & VISION 2000
Range	MULTI	Sr.No	210327882
Resolution	MULTI	Identification No	NA
DUC Condition	SATISFACTORY	Location	LABORATORY

Environmental Conditions & Calibration Procedure Details

Environmental Details	Temperature: 26.1° C	Relative Humidity	56% RH
Calibration Procedure No	SBS/CP/MD/19	Calibration done at	ONSITE

Reference Standards Details

S.No	Description	Make/ SI No:	Certificate No	Validity
1	Digital Light Meter	Lutron & S 034042	JRPM-CCTR-QPT-2022-0100	12-10-2023

CALIBRATION RESULTS

S.No.	DEVICE UNDER CALIBRATION READINGS	STANDARD INSTRUMENTS READINGS	DEVIATION	EXPANDED UNCERTAINTY (±)
LIGHT INTENSITY				
1	LUX 1100	LUX 1098	LUX 2	% 0.7

REMARKS

1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
2. The user should determine the suitability of the instrument for its intended use.
3. The recalibration interval should be determined on the user requirement.
4. The results stated in this certificate relate only to the item calibrated.
5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

Authorized Signatory

K.SH
 (Calibration Engineer)
 K.SATHYAMOORTHY

C.SR Technical Manager
 Chief Executive
 (C.SHANMUGARAJ)

