

CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/08818

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE,
POOVANTHI-630611,SIVAGANGAI DISTRICT.

SRF No.	SRF/23/00264-0054
SRF Date	18-07-2023
Date of Receipt	18-07-2023
Date of Calibration	18-07-2023
Due Date for Calibration	17-07-2024
Issue Date	20-07-2023

Details of Unit Under Calibration

Description	MICRO PIPETTE	Make	MICROPET
Range	5-50µl	Model	NA
Resolution	1µl	Material	PVC
Serial Number	NA	Operating Range	10-50µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition

Temperature (°C)	25.7	Humidity (%RH)	53
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6

Calibration Method Used

National / International Standard	ISO 8655-6:2002
Cal Procedure No	SBS/CP/ML/08

Standard Used

Sl. No.	Description	ID.No. / Sl. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSP22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319


Result of Calibration in µl

Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
1	10	9.89	9.88	9.87	9.88	9.88	9.88	-0.12	0.01	0.47
		9.88	9.86	9.87	9.88	9.89				
2	30	29.76	29.77	29.78	29.79	29.76	29.78	-0.22	0.01	0.47
		29.77	29.78	29.79	29.78	29.78				
3	50	49.88	49.87	49.88	49.87	49.88	49.87	-0.13	0.01	0.47
		49.87	49.86	49.86	49.87	49.88				

Remarks


- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- Equipment used for Calibration were calibrated & traceable to National & International Standards
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
- Calibration Liquid Used: Distilled or Deionized water conforming Grade 3 as specified in ISO 3696.

Calibrated By,


(Calibration Engineer)
M.RAGUL



Authorised by:


(Quality Manager/Chief Executive)
D.VETRI SELVI