

## CALIBRATION CERTIFICATE

<b>CERTIFICATE NO: SBS/CL/23/10362</b>		<b>MEDICAL DEVICES</b>		<b>Page No:1 of 1</b>
Issue Date	18-08-2023			
SRF No & Date	SRF/23/00320-0001 & 17-08-2023			
Receipt Date	17-08-2023			
Calibration Date	17-08-2023			
Calibration Due	16-08-2024			
<b>Customer Name &amp; Address</b>				
GOVERNMENT COMMUNITY HEALTH CENTRE, PUDUPETTAI- 635651, TIRUPATTUR DISTRICT.				
<b>Details of Device Under Calibration (DUC)</b>				
Description	: SEMI AUTO ANALYZER	Make & Model	: ROBONIK & PRIETEST TOUCH	
Range	: MULTI	Sr. No	: APCD2930186BU	
Resolution	: MULTI	Identification No	: NA	
DUC Condition	: SATISFACTORY	Location	: LABORATORY	
<b>Environmental Conditions &amp; Calibration Procedure Details</b>				
Environmental Details	Temperature:25.6 ° C	Relative Humidity	52% RH	
Calibration Procedure No	SBS/CP/MD/20	Calibration done at	ONSITE	
<b>Reference Standards Details</b>				
S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

### ELECTRICAL SAFETY

#### RESULTS

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resisitance	Measured values in MΩ	Uncertainty in % ( ± )
	>20MΩ	96	13.92
2	Earth Leakage	Measured values in μA	Uncertainty in % ( ± )
	<5000μAfor B,BF,CF	159	7.6
3	Enclosure Leakage	Measured values in μA	Uncertainty in % ( ± )
	<500μAfor B,BF,CF	205	7.3

#### REMARKS

- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

Authorised Signatory

(Calibration Engineer)  
M.RAGUL



Technical Manager  
C.SHANMUGARAJ

Chief Executive

## SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,  
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.