

rd Testing of Electrical, Electronic & Precision Instruments)

par. Co-operative Society, Besides Tadband Hanuman Temple, Sikh Village, Secunderabad - 500 009, T.S. INDIA ☎: (040) 27840697, 48541008  
 33008, 9948294008 Email: msudheer\_98@yahoo.com, msudheer\_98@smentp.org, SME\_003@YAHOO.IN. url : www.smentp.org


**CALIBRATION CERTIFICATE**

CALIBRATION ON 25/05/2023	CALIBRATION DUE 24/05/2024
1. Calibration Report Ref. No. & Date :	SME/NP/SVMC/DTHM/CAL-05/2023;Dt:25/05/2023
2. Requested by :	To, Prof & HOD (State reference Lab) Department Of Microbiology SV Medical College Tirupathi AP
3. Instrument Name	Digital Thermo Hygrometer
4. Make/Model:	HTC-2
5. Instrument ID No:	SVMC/DTHM/05
6. Range :	-10.0°C to 50.0 °C & 10% to 99 % RH
7. Accuracy :	+/- 1° C & 5%RH
8. Environmental Conditions :	Temp/ RH : 27 +/- 2Deg C; R.H:- 60 +/- 5%
9. Test parameter & condition :	Temperature & RH V/S Temperature& RH comprasion method
10. Standard Instruments used for Calibration:	Digital Thermo Hygrometer
11. Traceability Certificate Ref. No.	TSC/22-23/8627-1,Date:06/09/2022, Due Date:06/09/2023.

**TEST RESULTS**

SI No	Standard Reading Of Temperature & RH		Observed Reading Of Under Calibration Hygrometer In Deg C & %RH		Error Value +/-	
	Temperature in °C	RH in %	Temperature in °C	RH in %	Temp in °C	RH in %
1	-10.0	10	-10.1	11	0.1	1
2	0.0	30	0.2	31	0.2	1
3	10.0	40	10.2	41	0.2	1
4	20.0	50	20.2	50	0.2	0
5	30.0	60	30.3	62	0.3	2
6	40.0	75	40.1	76	0.1	1
7	50.0	90	49.9	87	0.1	3

PLEASE NOTE: THE INSTRUMENTS ERROR NOT FOUND TO BE WITHIN THE SPECIFIED ACCURACY LIMIT ANDTHIS CERTIFICATE VALID FOR PERIOD OF ONE YEAR FROM THE DATE OF ISSUE.

CALIBRATED BY  
 S.M. ENTERPRISES  
  
 S.M. ENTERPRISES  
 QA Incharge  
 ENGINEER INCHARGE  
 Reg. No. RQ9113  
 SECUNDERABAD