

CALIBRATION CERTIFICATE

CALIBRATION ON 25/05/2023	CALIBRATION DUE 24/05/2024
1. Calibration Report Ref. No. & Date :	SME/NP/SVMC/TIM/05/2023;Dt:25/05/2023
2. Requested by : (Name and address of the organization)	To, Prof & HOD (State reference Lab) Department Of Microbiology SV Medical College Tirupathi AP
3. Instrument Name:	Digital Timer
4. Instrument Make/Model:	Aceteq / DT-1045
5. Instrument ID. No:	SVMC/TIM/05
6. Range :	0 to 99 min
7. Accuracy :	+/- 1% FSD
8. Environmental Condition	Temp/RH : 28 +/- 2 Deg C; R.H: 65 +/- 5%
9. Test parameter & condition :	Time V/S Time (Comparision Method)
10. Standard Instruments used for Calibration:	Time Interval Meter
11. Traceability Certificate Ref. No.	C-220622-5-1 ,Date: 29/06/2022 , Due Dt: 29/06/2023

TEST RESULTS

Sl No	Standard Reading Of Time Hr:min:sec:ms	Observed Reading Of Time min:sec	Error Value(+/-) Hr:min:sec:ms
1	00:01:00:29	01:00	00:00:00:29
2	00:10:00:18	10:00	00:00:00:18
3	00:30:00:29	30:00	00:00:00:29
4	01:50:00:32	90:00	00:00:00:32

PLEASE NOTE: THE INSTRUMENTS ERROR FOUND TO BE WITHIN THE SPECIFIED ACCURACY LIMIT ANDTHIS CERTIFICATE VALID FOR PERIOD OF ONE YEAR FROM THE DATE OF ISSUE.

