

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME





Lab Name DISTRICT HOSPITAL, VIDISHA Lab No 7823

Constituent HbA1c Date of Result Entered: 17/06/2023 Group

PT item Date of Report Published : 04/07/2023 Lyophilized human whole blood based

| 1 HbA1c HPLC-lon Exchange Biorad (D10 / D100 / Variant Turbo) 715 6.80 2.65 0.18 6.8 % 0.00 0.01 | SI.N | o Analyte | Method / Principle Name | Analyzer N | er Name | No of Participants | DV | Participants | | Your Value | SDI | U |
|--|---------------------------------------|-----------|----------------------------|--------------|----------------|-----------------------|------|--------------|------|---------------|------|------|
| - Variant Turbo) | I I I I I I I I I I I I I I I I I I I | | Biorad ([| D10 / D100 / | | ا مما | CV | SD | 1 1 | 0.00 | | |
| SDI Range Interpretation | | Val | | | t Turbo) | 715 | 0.00 | 2.65 | 0.10 | 0.0 70 | 0.00 | 0.01 |
| · | SDI Range | | | | Interpretation | | | | | | | |

| SDI Range | Interpretation | | | |
|-----------------------|--|--|--|--|
| Within -1.00 to +1.00 | Excellent. | | | |
| Within ±1.01 to ±2.00 | Good. | | | |
| Within ±2.01 to ±2.99 | Accept with caution. Warning Signal. | | | |
| Beyond ±3.0 | Unacceptable performance. Action Signal. | | | |

LAB ADDRESS:

DISTRICT HOSPITAL, VIDISHA DISTRICT HOSPITAL, VIDISHA **VIDISHA MADHYA PRADESH464001**

> **Coordinator Contact Details:** Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102

Panela Christudoss Dr. Pamela Christudoss **CMC EQAS Coordinator**

Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed. Data in CMC EQAS reports is confidential CMC EQAS does not sub contract any components ****** End of Report *****