


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



Calibration Certificate

| | | | | |
|--|----------------------------|--------------|-----------------|------------|
| CERTIFICATE NO. : 230630B/N-116 | Page : 1 of 1 | | | |
| CALIBRATED ON. : 05 / 07 / 2023 | Ref. Std. : CG-11 | | | |
| CALIBRATION DUE. : 04 / 07 / 2024 | C.P. No. : ACS/CP/07 | | | |
| SRF No. & DATE : 230630B/N Dt. 30.06.2023 | Phy. Inspection : O.K. | | | |
| ULR No. & ISSUE DATE : CC342723000000709F Dt. 07.07.2023 | Cal. At : Site / Lab | | | |
| CALIBRATED FOR : M/s Redcliffe Lifetech Pvt. Ltd. 1 st Floor, Shop No.1 Nile complex Old Hazaribagh Road Kantatoli Ranchi Jharkhand-834001 | | | | |
| ITEM CALIBRATED : Digital Temperature Indicator | | | | |
| Make : RoHS | Range : (-50 to 110) °C | | | |
| L. Count : | ID No. : RCL/HEM/D.Temp-04 | | | |
| Location : Haematology | S. No. : N.A | | | |
| ENVIRONMENT CONDITION: Temp. { 25°C ± 3°C} R.H. {40% to 60%} | | | | |
| STANDARD INSTRUMENTS USED : | | | | |
| INSTT. NAME | ID No. / S. No. | TRACEABILITY | CERTIFICATE No. | DUE DATE |
| Digital Calibrator | ACS/UC/01 / 300561 | Anshaankan | AIPL/2022/0846 | 10/11/2023 |
| RESULTS : ET Discipline | | | | |
| * U. C. INSTT. READS (°C) | STANDARD READING (°C) | ERROR (°C) | | |
| 2.08 | 2 | 0.08 | | |
| 4.16 | 4 | 0.16 | | |
| 6.20 | 6 | 0.20 | | |
| 8.26 | 8 | 0.26 | | |
| (average of readings) | | | | |

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 1.00°C at coverage Factor k=1.96.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

Auth. Signatory
S.K. Sharma
CEO/TM