


# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



## Calibration Certificate

<b>CERTIFICATE NO.</b>	: 230630B/N-108	<b>Page</b>	: 1 of 1	
<b>CALIBRATED ON.</b>	: 30 / 06 / 2023	<b>Ref. Std.</b>	: DKD R-5-7	
<b>CALIBRATION DUE.</b>	: 29 / 06 / 2024	<b>C.P. No.</b>	: ACS/CP/11	
<b>SRF No. &amp; DATE</b>	: 230630B/N Dt. 30.06.2023	<b>Phy. Inspection</b>	: O.K.	
<b>ULR No. &amp; ISSUE DATE</b>	: CC342723000000701F Dt. 07.07.2023	<b>Cal. At</b>	: Site / Lab	
<b>CALIBRATED FOR</b>	: M/s Redcliffe Lifetech Pvt. Ltd. 1 <sup>st</sup> Floor, Shop No.1 Nile complex Old Hazaribagh Road Kantatoli Ranchi Jharkhand-834001			
<b>ITEM CALIBRATED</b>	: Digital Thermo Hygrometer			
	<b>Make</b>	: Digital Series	<b>Range</b>	: (0 to 50)°C, (10-99)%
	<b>L. Count</b>	:	<b>ID No.</b>	: RCL/BIO/HGM-02
	<b>Location</b>	: Clinical Biochemistry	<b>S. No.</b>	: N.A
<b>ENVIRONMENT CONDITION:</b>	Temp. { 25°C ± 3°C } R.H. {40% to 60%}			
<b>STANDARD INSTRUMENTS USED :</b>				
<b>INSTT. NAME</b>	<b>ID No. / S. No.</b>	<b>TRACEABILITY</b>	<b>CERTIFICATE No.</b>	<b>DUE DATE</b>
RH & Temp. Senor with Indicator	ACS/RHTIS/01/ 20365872	DVG	CC210122000001272F	16/07/2023
<b>RESULTS</b>	: Thermal Discipline			
<b>* U. C. INSTT. READS</b>	<b>STANDARD READING</b>	<b>ERROR</b>	<b>UNC(±)</b>	
(°C)	15.7	14.91	0.79	2.1
	25.4	24.92	0.48	2.1
	40.5	39.93	0.57	2.1
(average of readings)				
(%)	30.4	29.95	0.45	2.5
	60.5	59.96	0.54	2.5
	90.6	89.93	0.67	2.5
(average of readings)				

### REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 2.1°C, 2.5% at coverage Factor k=1.96.

\*U.C.

: Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

