


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO. : 230630B/N-102	Page : 1 of 1			
CALIBRATED ON. : 30 / 06 / 2023	Ref. Std. : ISO:8655:6			
CALIBRATION DUE. : 29 / 06 / 2024	C.P. No. : ACS/CP/08			
SRF No. & DATE : 230630B/N Dt. 30.06.2023	Phy. Inspection : O.K.			
ULR No. & ISSUE DATE : CC342723000000695F Dt. 07.07.2023	Cal. At : Site / Lab			
CALIBRATED FOR : M/s Redcliffe Lifetech Pvt. Ltd. 1 st Floor, Shop No.1 Nile complex Old Hazaribagh Road Kantatoli Ranchi Jharkhand-834001				
ITEM CALIBRATED : Micro Pipette				
Make : Thermo Scientific	Range : (10 to 100) µl			
L. Count :	ID No. : RCL/HEM/PIP-02			
Location : Haematology	S. No. : TW06024			
ENVIRONMENT CONDITION: Temp. { 25°C ± 3°C } R.H. {40% to 60%}				
STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/D449927064	BlueBox	BBT/398/JAN/23	05/01/2024
RESULTS : Mechanical Discipline				
Nominal Capacity (µl)	Graduation (µl)	Nominal Value (µl)	Volume at 20 °C (µl)	Error (µl)
		10	10.10	(-) 0.10
100		50	50.17	(-) 0.17
		100	100.39	(-) 0.39

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 0.6 µl at coverage factor k=2.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

