

Erba Lachema s.r.o.




Brno, 10.11.2016

To,  
Redcliffe Lifetech Private Limited,  
1<sup>st</sup> Floor Shop No.1, Nile Complex Old Hazaribagh Road Kantatoli  
Ranchi Jharkhand - 834001

Instrument Name: ERBA LAURA  
Instrument Serial No: 1019-1-202206

We Erba Lachema SRO with registered office in Czech Republic, Karasek 2219/1D, 621 00 BMO, represented by its Production manager Veronika Pleskova here by confirm as Producer of LAURA instrument that this device is firmly calibrated by the manufacturer and does not require any regular calibration by the user, correct calibration of instrument is controlled by grey strips provided initially together with the instrument.

*R.P.*  
  
RNDr. Pavel Suchan Ph.D.

**Erba Lachema s.r.o.**  
Karásek 1d, 621 00 Brno  
DIČ: CZ26918846 -10-

Global Product Manager, Urine Analysis

Erba Lachema s.r.o.

Karásek 1d, Brno 62100, Czech Republic



Erba Lachema s.r.o., Karásek 1d, 621 00 Brno, Czech Republic  
Identification number: 269 18 846. Tax identification number: CZ26918846  
Incorporated in the Commercial Register maintained by the Regional Court in Brno, Section C, insert 45458  
Tel.: +420 517 077 111. fax: +420 517 077 077. e-mail: [diagnostics@erbalachema.com](mailto:diagnostics@erbalachema.com). [www.erbalachema.com](http://www.erbalachema.com)

\*\*\*\*\*

**Erba** Mannheim Urinalysis

LAURA  
\*\*\*\*\*  
09.19.2023 08:23

DEKAPHAN LAURA  
Seq.No: 0001  
URINORM P  
09.19.2023 08:27

*BLD	10	Ery/u1
*BIL	3	mg/dl
*UBG	1	mg/dl
*KET	52	mg/dl
*GLU	1000	mg/dl
*PRO	100	mg/dl
pH	9	
*NIT	POS	
*LEU	TRACE	
SG	1.000	

-----  
QC TEST: OK

DEKAPHAN LAURA  
Seq.No: 0002  
URINORM N  
09.19.2023 08:32

BLD	NEG
BIL	NEG
UBG	NORM
KET	NEG
GLU	NEG
PRO	NEG
pH	6
NIT	NEG
LEU	NEG
SG	1.030

-----  
QC TEST: OK

*Sumadi*  
19/9/23