



BIO-RAD LABORATORIES (INDIA) PRIVATE LIMITED

Site name:	DISTRICT HOSPITAL
Address:	Sanchi Rd , Vidisha, Madhya Pradesh 464001
Province-Postal code	464001
Contact Name	Mr. Abhishek Prajapati

Calibration Certificate

Date	23.11.2022	Instrument Model	D-10 Hb Analyzer	Instrument Serial Number	DJ0F517660
------	------------	------------------	------------------	--------------------------	------------

We certify that the Bio-Rad instrument specified above and installed in your lab is calibrated and the performance was found satisfactory.

High Pressure Pump Flow Rate Reference: Service Manual: Section 7.7.3

Measured with (Device manufacturer and model)	Device Serial Number	Device Calibration date (dd mmmm, yyyy)
Measuring cylinder hex base 10ML	2310100	09.09.22
Specified Range@ 1.5ml/Minute	Measured Value	
1.425 – 1.575 ml/minute	1.49	

Column Holder temperature Reference: Service Manual: Section 7.9

Measured with (Device manufacturer and model)	Device Serial Number	Device Calibration date (dd mmmm, yyyy)
FLUKE 179	26030078	06.06.22
Specified Range @ 35°C	Measured Value	
33.4°C – 36.4°C	34.4	

Detector log value Reference: Service Manual: Section 7.2 or Field Bulletin D-10-081

Measured with	Specified Range	Measured Value
Instrument Service Software	0.018 – 0.022 AU	0.021AU

The next calibration is due on	22.11.2023	(dd mmmm, yyyy)
--------------------------------	------------	-----------------

For Bio-Rad Laboratories India Pvt Limited

Authorized Signatory:

Sarav Penth