



LAB : Plot No. 272-C, Harbanspura, Near Govt. I.T.I., Yamuna Nagar - 135001, Haryana.
Phone 01732-252120, 9991584585, 9050407120, E-mail : microtechnology.ind@gmail.com

CALIBRATION CERTIFICATE

Customer Name And Address	M/S : Shrimant Madhavrao Scindia District Govt. Hospital Vidisha Central Processing Lab Sanchi Road Vidisha, MP – 464001.	Certificate No.	MT/09/150923/MC7/001768
		ULR No.	CC333323000001767F
		SRF No. & Date	MT/09/15.09.23
		Receipt Date	15/09/2023
		Date of Calibration	15/09/2023
		Suggested Due Date	14/09/2024
		Date of Issue	16/09/2023

Instrument Details			
Instrument name	Dig. Temp. Controller with Sensor (Refrigerator)	Sr. No.	----
Make / Model No.	Videocon	Location	Lab
Range	2° to 8° C	Accuracy	-----
Least Count	1 °C	Visual Inspection	OK
I.D. No.	DH VIDISHA/CPL/REF/14	Calibration Performed At	Site

Detail of Reference Standard & Major Equipment Used			
Equipment Name	4 Wire RTD With Indicator		
Make	Tempsens/ Yudian		
Model / SR No.	1360/ 19F330198		
Certificate No.	TL/022/1156.2.1		
Calibration Validity	16.10.2023		
Calibration by	Tempsens Calibration Centre		

Environmental Condition	Temperature	25±4 °C	Calibration Reference	IS:DKD-R 5-7
	Relative Humidity	50±20 %	Calibration Procedure	CP/T/03

Calibration Results			
Serial No.	UUC Value in (°C)	Std. (Master) Value in (°C)	Uncertainty At 95% C.L. (coverage factor K=2)
01.	2	2.16	±0.71 °C
02.	4	4.27	±0.71 °C
03.	6	6.39	±0.71 °C
04.	8	8.52	±0.71 °C

Remarks:

- ❖ (1) Standard equipment use for calibration are traceable to national/ international standards.
- ❖ (2) The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor $k = 2$ such that the coverage probability corresponds to approximately 95%.
- ❖ (3) The above results are valid at the time of and under the stated conditions measurement.
- ❖ (4) This certificate is refers only to the particular item submitted for calibration.
- ❖ (5) Next calibration due date given as requested by the customer.

Calibrated By
(Calibration Engineer)
(Ankush Kumar)
Format No. F01(7.8)



.....End of the Certificate.....

Approved By
(Quality Manager)
(Amit Sajji)
Page No. 1 of 1