



LAB : Plot No. 272-C, Harbanspura, Near Govt. I.T.I., Yamuna Nagar - 135001, Haryana.
Phone 01732-252120, 9991584585, 9050407120, E-mail : microtechnology.ind@gmail.com

CALIBRATION CERTIFICATE

Customer Name And Address	M/S : Shrimant Madhavrao Scindia District Govt. Hospital Vidisha Central Processing Lab Sanchi Road Vidisha, MP – 464001.	Certificate No.	MT/09/150923/MC7/001772
		ULR No.	CC333323000001771F
		SRF No. & Date	MT/09/15.09.23
		Receipt Date	15/09/2023
		Date of Calibration	15/09/2023
		Suggested Due Date	14/09/2024
		Date of Issue	16/09/2023

Instrument Details			
Instrument name	RPM Measurement (Centrifuge M/C)	S. No.	B.G.V.
Make	Cell Win	Location	Lab
Range / Size	0 - 3000 RPM	Accuracy	-----
Least Count	As Per range	Visual Inspection	OK
I.D. No.	DH VIDISHA/CPL /CENTRIFUGE/20	Calibration Performed At	Site

Detail of Reference Standard & Major Equipment Used			
Equipment Name	Digital Tachometer		
Make	Line Seiki		
Model / SR No.	164-0142		
Certificate No.	40099763		
Calibration Validity	06/12/2023		
Calibration by	Belz calibration Lab		

Environmental Condition	Temperature	23±5° C	Calibration Reference	NABL 129 , 130
	Relative Humidity	50±20 %	Calibration Procedure	CP/M/A&S/01

Calibration Results

Sr. No.	UUC Value In (rpm)	STD (Master) Value In (rpm)	Uncertainty At 95% C.L. (coverage factor k=2)
01.	1000	999.6	± 5.6%rdg.
02.	1500	1498.3	± 5.6%rdg.
03.	2000	1995.8	± 5.6%rdg.
04.	3000	2993.4	± 5.6%rdg.

Remarks:

- ❖ (1) Standard equipment use for calibration are traceable to national/ international standards.
- ❖ (2) The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor $k = 2$ such that the coverage probability corresponds to approximately 95%.
- ❖ (3) The above results are valid at the time of and under the stated conditions measurement.
- ❖ (4) This certificate is refers only to the particular item submitted for calibration.
- ❖ (5) Next calibration due date given as requested by the customer.

Calibrated By
(Calibration Engineer)
(Kanhaiya)
Format No. F01(7.8)



.....End of the Certificate.....

Approved By
(Quality Manager)
(Amit Saini)
Page No. 1 of 1