



UNIQUE CALIBRATION & CONSULTANT SERVICES

AN ISO 9001:2015 CERTIFIED CALIBRATION LABORATORY



CC-3287

CALIBRATION CERTIFICATE

Format No.: UCCS-QF-7.8.1		Certificate No.: UCCSCC20230911-009		ULR No.: CC328723000008449F			
Group: Temperature & Humidity			Discipline: Thermal				
Customer's Name & Address M/s. SHRIMANT MADHAVRAO SCINDIA DISTRICT GOVT. HOSPITAL VIDISHA CENTRAL PROCESSING LAB, SANCHI ROAD, VIDISHA, MP -464001			Page No.: 1 of 1	Calibration Performed: At Lab			
			Date of Receipt: 11-09-2023			Service Request No.: UCCSSRN20230911-004	
			Certificate Issue Date: 13-09-2023			Date of Calibration: 12-09-2023	
			Next Due Date (Suggested by Customer): 11-09-2024				
			Equipment Description (UUC) & Identification			Name of the Instrument: Digital Thermo Hygrometer	
			Make: HTC-2		Model: -----		
Range: Up to 50 °C & Up to 90 %RH			Sr. No.: -----		I.D. No.: DH VIDISHA/CPL/HM/19		
Least Count: 0.1 °C & 1 %RH			Location: Lab				
Condition of Instrument: Ok			Master Equipment Used				
Sr. No.	Instrument's Name	Range & I.D. No.	Calibration Agency	Calibration Certificate No.	Valid Up to		
1	Dig. Temp. & Hum. Meter With Ext. Sensor	0 °C to 50 °C , 0 to 100 % RH & UCCS/CAL/030	Acumen Measurements & Consultancy Pvt. Ltd.	AMC/230207/01	03-08-2024		
2	High Precision Thermometer with 4 Wire RTD Sensor	-100 °C to 600 °C & UCCS/CAL/021	Bansal Calibration Services	BCS/N/23030005-001	02-03-2024		
Environmental Conditions			Calibration Procedure	Calibration Reference Standard			
Temperature: 25 ± 4 °C, Humidity: 50 ± 10 %RH			UCCS 03 SOP 024 & 028	-----			

CALIBRATION RESULTS

Digital Temperature Indicator With Inbuilt Sensor					
Sr. No.	Set Value (°C) @ 50 %RH	UUC Value (°C)	Ref. Standard Value (°C)	Error (°C)	Expanded Uncertainty (± °C)
1	15	15.1	14.98	0.12	0.5
2	25	25.1	24.91	0.19	0.5
3	30	30.0	29.77	0.23	0.5
4	45	45.2	44.93	0.27	0.5

Sr. No.	Set Value (%RH)@ 25 °C	UUC Value (%RH)	Ref. Standard Value (%RH)	Error (%RH)	Expanded Uncertainty (± %RH)
1	30	30	30.3	-0.3	2.33
2	45	45	45.7	-0.7	2.33
3	60	60	60.9	-0.9	2.33
4	90	90	91.1	-1.1	2.33

End of Certificate

Calibrated By
Amit Kumar

Calibration Engineer

Reviewed/Approved By

L.M. Mahato
Chief Executive Officer (CEO)

Note: - 1. The calibration results reported in this certificate are valid at the time of and stated condition of measurement.

2. This report is not to be reproducing wholly or in part and cannot be used as evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
3. Total liability of our Organization is limited to the invoiced amount.
4. Calibration Certificate without signature are not valid.
5. All our certificates are metrological traceable.
6. The results relate only to the item calibrated.
7. UUC: Unit Under Calibration.
8. Expanded Uncertainty at 95 % confidence level with coverage factor k=2.
9. Scan this QR code to download our NABL Accreditation Scope & Certificate.



Working : W.Z.1390/Z-2, 1st Floor, Nangal Raya Extn., Pankha Road, New Delhi-110046

Regd. Office : R.Z.56A, Indra Park, Gali No.25, Palam Colony, New Delhi-110045

M.: 9958445875, 9313768426, 9999761885 E-Mail : uniquecalibrationservices@gmail.com