

# CALIBRATION REPORT

Centrifuge Table top



Report No:	CAL-AR-2023-02-1256	Date:20-02-2023	Page 01 of 02
------------	---------------------	-----------------	---------------

Customer Details			
Health Facility	Aalo GH		
State	Arunachal Pradesh	District	West Siang

Medical Equipment Details			
ME Number	ARZ1WESDH0130016	Manufacturer	Alamicro
ME Name	Centrifuges, Tabletop (Centrifuge)	Model	NA
Department	Ictc-Na	Serial No	NA
Calibrated at	On Site	Condition of ME	Good
Date of Calibration	11-02-2023	Due for Calibration	10-02-2024

Environmental Conditions			
Temperature in °C	20°C	Humidity in % Rh	58%

Details of Master Equipment Used for Calibration						
No	Equipment ID	Description	Manufacturer	Model	Serial No.	Calibration Validity
1	TRI-TE-0380	Digital Thermometer	METRAVI	DTM-900	210654960	07-05-2023

Performance Results						
No	Parameters	Unit	Set Value	Measured Value	Acceptable Range	Remarks
1	Speed	rpm	1500	1520	1425 to 1575	Pass
2	Speed	rpm	2500	2530	2375 to 2625	Pass
3	Speed	rpm	3500	3550	3325 to 3675	Pass

# CALIBRATION REPORT

Centrifuge Table top



## Conditions of Issue of report

1. The Standards used are traceable to National standards.
2. This report should not be reproduced except in full without written approval.

Calibration Status

Passed

Failed

Calibrated By

A handwritten signature in black ink, appearing to be "Balakyrshanskhem Kurbah".

Balakyrshanskhem Kurbah

Authorised Signatory

A handwritten signature in black ink, appearing to be "Rahul Kumar Kashyap".

Rahul Kumar Kashyap