

CALIBRATION REPORT

Refrigerators, Laboratory(Lab refrigerator)



Report No:	CAL-AR-2023-02-1419	Date: 04-03-2023	Page 01 of 02
------------	---------------------	------------------	---------------

Customer Details			
Health Facility	Aalo GH		
State	Arunachal Pradesh	District	West Siang

Medical Equipment Details			
ME Number	ARZ1WESDH0130315	Manufacturer	Godrej
ME Name	Refrigerators, Laboratory (Lab refrigerator)	Model	Edge
Department	Laboratory	Serial No	NA
Calibrated at	On Site	Condition of ME	Good
Date of Calibration	25-02-2023	Due for Calibration	24-02-2024

Environmental Conditions			
Temperature in °C	20°C	Humidity in % Rh	58%

Details of Master Equipment Used for Calibration						
No	Equipment ID	Description	Manufacturer	Model	Serial No.	Calibration Validity
1	TRI-TE-0380	Digital Thermometer	METRAVI	DTM-900	210654960	07-05-2023

Performance Results						
No	Parameter	Set Temp	Unit	Measured Value	Acceptable Range	Remarks
1	Temperature	2°C	°C	2.1	1.5 to 2.5°C	Pass
2	Temperature	4°C	°C	4.2	3.5 to 4.5°C	Pass
3	Temperature	6°C	°C	6.3	5.5 to 6.5°C	Pass

CALIBRATION REPORT

Refrigerators, Laboratory(Lab refrigerator)



Conditions of Issue of report

1. The Standards used are traceable to National standards.
2. This report should not be reproduced except in full without written approval from Trimed Solutions India Private Limited.

Calibration Status

Passed

Failed

Calibrated By

A handwritten signature in blue ink, appearing to read "T. Zirido".

Tomin Zirido

Authorised Signatory

A handwritten signature in blue ink, appearing to read "Rahul Kumar Kashyap".

Rahul Kumar Kashyap