



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JULY 2023



Lab Name	Govt. Block PHC, Mallasamudram	Lab No	7447
Constituent Group	Chemistry III	Date of Result Entered :	20/07/2023
PT item	Lyophilized human serum based	Date of Report Published :	07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	106 mg/dL	-0.55	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	6 mg/dL	0.49	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	114 mg/dL	0.21	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. Block PHC, Mallasamudram
 THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, NAMAKKAL DISTRICT
 Mallasamudram
 TAMILNADU637503

Coordinator Contact Details:
 Email:clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential