



# S.M. ENTERPRISES

# 27<sup>th</sup>

ISO 9001 : 2015  
RQ91/3144

(Calibration and Testing of Electrical, Electronic & Precision Instruments)

Years of Service

Office/Cal. Lab : Plot No. 12, "Ganesh Niwas", Sanjeevaiah Nagar, Co-operative Society, Besides Tadbund Hanuman Temple, Sikh Village, Secunderabad - 500 009, T.S, INDIA ☎: (040) 27840697, 48541008  
Cell: 9848020798, 9948291008, 9948292008, 9848293008, 9948294008 Email: msudheer\_98@yahoo.com, msudheer\_98@smentp.org, SME\_003@YAHOO.IN, url : www.smentp.org

## CALIBRATION CERTIFICATE

CALIBRATION ON 15/05/2023	CALIBRATION DUE 14/05/2024
1. Calibration Report Ref. No. & Date :	SME/NP/GMC/DTHM/027/2023;Dt:15/05/2023
2. Requested by : (Name and address of the organization)	To, Prof & HOD( Department Of Microbiology). Government Medical College,Ananthapur.
3. Instrument Name	Digital Thermo Hygrometer
4. Make/Model:	HTC-2
5. Instrument ID No:	GMC/DTHM/027
6. Range :	-10.0°C to 50.0 ° C & 10% to 99 % RH
7. Accuracy :	+/- 1° C & 5%RH
8. Environmental Conditions :	Temp/ RH : 27 +/- 2Deg C; R.H:- 60 +/- 5 %
9. Test parameter & condition :	Temperature & RH V/S Temperature& RH comprasion method
10. Standard Instruments used for Calibration:	Digital Thermo Hygrometer
11. Traceability Certificate Ref. No.	TSC/22-23/8627-1,Date:06/09/2022,Due Date:06/09/2023.

TEST RESULTS							
SI No	Standard Reading Of Temperature & RH		Observed Reading Of Under Calibration Hygrometer In Deg C & %RH		Error Value +/-		
	Temperature in °C	RH in %	Temperature in °C	RH in %	Temp in °C	RH in %	
1	-10.0	10	-10.0	11	0.0	1	
2	0.0	30	0.3	32	0.3	2	
3	10.0	40	10.5	43	0.5	3	
4	20.0	50	20.2	53	0.2	3	
5	30.0	60	30.3	63	0.3	3	
6	40.0	75	40.5	78	0.5	3	
7	50.0	90	49.7	88	0.3	2	

PLEASE NOTE: THE INSTRUMENTS ERROR NOT FOUND TO BE WITHIN THE SPECIFIED ACCURACY LIMIT ANDTHIS CERTIFICATE VALID FOR PERIOD OF ONE YEAR-FROM THE DATE OF ISSUE.

CALIBRATED BY  
S.M.ENTERPRISES

