

CENTRIFUGE Calibration / Verification Record Sheet

Location: ICTC-CHC-GOOTY

Range : Up to 3500 RPM

Equipment ID: APSACS/ICTC/CHC/GTY/CEF/01

SL. No:2

Calibration Procedure Reference: In house verification of centrifuge using standard calibrated by Tachometer

Date of Calibration	20-09-2023	Due Date of Calibration	19-09-2024
Certified Centrifuge ID: GMC-ATP/HIVTL/SRL/CEF/01	Tachometer ID: GMC-ATP/HIVTL/SRL/TM/01	Calibration date: 19.05.2023	Due date of Calibration: 18.05.2023
Verification Done By: P. ANURADHA, Technical Officer, SRL, GMC, ANANTHAPURAMU			
No.	STANDARD READI NG (RPM)	UUC READING (RPM)	ERROR (RPM)
1.	499.7	500	0.3
2.	999.2	1000	0.8
3.	1998.7	2000	1.3
4.	2997.5	3000	2.5
5.	3496.8	3500	3.2

Accept/ Reject: Accepted since average difference is within +/-1000RPM

Action Taken: NA

Remarks: NA

P. Anuradha

SRL-1.0

Dept. of Microbiology

GMC - Anantha Puram.