



ISO 9001 : 2015
RQ91/3144

S.M. ENTERPRISES

(Calibration and Testing of Electrical, Electronic & Precision Instruments)

27th

Years of Service

Office/Cal. Lab : Plot No. 12, "Ganesh Niwas", Sanjeevaiah Nagar, Co-operative Society, Besides Tadbund Hanuman Temple, Sikh Village, Secunderabad - 500 009, T.S, INDIA ☎: (040) 27840697, 48541008
Cell: 9848020798, 9948291008, 9948292008, 9848293008, 9948294008 Email: msudheer_98@yahoo.com, msudheer_98@smentp.org, SME_003@YAHOO.IN, url : www.smentp.org

CALIBRATION CERTIFICATE

| CALIBRATION ON 15/05/2023 | CALIBRATION DUE 14/05/2024 |
|---|--|
| 1. Calibration Report Ref. No. & Date : | SME/NP/SMC/TIM/03/2023;Dt:15/05/2023 |
| 2. Requested by : (Name and address of the organization) | To, Prof & HOD(Department Of Microbiology). Siddhartha Medical College,Vijayawada. |
| 3. Instrument Name: | Digital Timer |
| 4. Instrument Make/Model: | ACETEQ / DT-1045 |
| 5. Instrument ID. No: | SMC/TIM/03 |
| 6. Range : | 0 to 99 min |
| 7. Accuracy : | +/- 1 % FSD |
| 8. Environmental Condition | Temp/RH : 28 +/- 2 Deg C; R.H: 65 +/- 5% |
| 9. Test parameter & condition : | Time V/S Time (Comparison Method) |
| 10. Standard Instruments used for Calibration: | Time Interval Meter |
| 11. Traceability Certificate Ref. No. | C-220622-5-1 ,Date: 29/06/2022 , Due Dt: 29/06/2023 |

TEST RESULTS

| SI No | Standard Reading Of Time Hr:min:sec:ms | Observed Reading Of Time min:sec | Error Value(+/-) Hr:min:sec:ms |
|-------|---|-------------------------------------|-----------------------------------|
| 1 | 00:01:00:39 | 01:00 | 00:00:00:39 |
| 2 | 00:10:00:50 | 10:00 | 00:00:00:50 |
| 3 | 00:30:00:45 | 30:00 | 00:00:00:45 |
| 4 | 01:50:00:37 | 90:00 | 00:00:00:37 |

PLEASE NOTE: THE INSTRUMENTS ERROR FOUND TO BE WITHIN THE SPECIFIED ACCURACY LIMIT AND THIS CERTIFICATE VALID FOR PERIOD OF ONE YEAR FROM THE DATE OF ISSUE.

CALIBRATED BY

S . M. Enterprises

ENGINEER IN CHARGE

