



S.M. ENTERPRISES

27th

ISO 9001 : 2015
RQ91/3144

(Calibration and Testing of Electrical, Electronic & Precision Instruments)

Years of Service

Office/Cal. Lab : Plot No. 12, "Ganesh Niwas", Sanjeevaiah Nagar, Co-operative Society, Besides Tadbund Hanuman Temple, Sikh Village, Secunderabad - 500 009, T.S, INDIA ☎: (040) 27840697, 48541008
Cell: 9848020798, 9948291008, 9948292008, 9848293008, 9948294008 Email: msudheer_98@yahoo.com, msudheer_98@smentp.org, SME_003@YAHOO.IN, url : www.smentp.org

CALIBRATION CERTIFICATE

CALIBRATION ON 15/05/2023	CALIBRATION DUE 14/05/2024
1. Calibration Report Ref. No. & Date :	SME/NP/SMC/DTHM/02/2023;Dt:15/05/2023
2. Requested by :	To, Prof & HOD(Department Of Microbiology). Siddhartha Medical College,Vijayawada.
3. Instrument Name	Digital Thermo Hygrometer
4. Make/Model:	HTC-2
5. Instrument ID No:	SMC/DTHM/02
6. Range :	-10.0°C to 50.0 °C & 10% to 99 % RH
7. Accuracy :	+/- 1°C & 5%RH
8. Environmental Conditions :	Temp/ RH : 27 +/- 2Deg C; R.H:- 60 +/- 5 %
9. Test parameter & condition :	Temperature & RH V/S Temperature& RH comprasion method
10. Standard Instruments used for Calibration:	Digital Thermo Hygrometer
11. Traceability Certificate Ref. No.	TSC/22-23/8627-1,Date:06/09/2022,Due Date;06/09/2023.

TEST RESULTS

Sl No	Standard Reading Of Temperature & RH		Observed Reading Of Under Calibration Hygrometer In Deg C & %RH		Error Value +/-	
	Temperature in °C	RH in %	Temperature in °C	RH in %	Temp in °C	RH in %
1	-10.0	10	-10.0	12	0.0	2
2	0.0	30	0.1	32	0.1	2
3	10.0	40	10.3	37	0.3	3
4	20.0	50	20.1	52	0.1	2
5	30.0	60	29.7	61	0.3	1
6	40.0	75	39.9	78	0.1	3
7	50.0	90	49.8	89	0.2	1

PLEASE NOTE: THE INSTRUMENTS ERROR NOT FOUND TO BE WITHIN THE SPECIFIED ACCURACY LIMIT ANDTHIS CERTIFICATE VALID FOR PERIOD OF ONE-YEAR FROM THE DATE OF ISSUE.

CALIBRATED BY
S.M.ENTERPRISES